

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 10 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130134</u> ✓
Name		<u>JSwartz</u>	ID# <u>4314</u> Date <u>08/28/09</u> ✓
A	Agency	<u>SCDL</u>	Phone # <u>269-5592</u>
Instrument Location		<u>SCDL</u>	✓
Alco S/N		<u>SCOTTUS</u> ✓	Target Value <u>.080</u> ✓ High Pressure <u>600 psi</u>
B	Alco Test Values	<u>.078</u> ✓ 1 st Alco	<u>.078</u> ✓ 2 nd Alco
Signature		<u>[Signature]</u> ✓	
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 8th day of Oct, 2009.

[Signature]

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130134 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Blank box for Nondrinking Subject Test tape.

DIAGNOSTIC CHECK

Blank box for Diagnostic Check tape.

Breath Test Record
Serial Number 130134 Model DataMaster
Location HNH1 Hoonah Police Dept. ✓

Subject Information
Last Name Verification Of Cal O.L. # L
First Name L

Operator Information
Operator Name Swartz/J Case
Operator # 4314 Department AAA1

Test Results
Alco Target .080 ✓ Alco Serial SCOTT08 ✓
Nominal 0.078 Barometer 29.47 Test Type V
Dry Gas 1 0.078 ✓ Time 10:44:03 Test Date 08/28/2009
Subject Alcohol 0.000 ✓ Time 10:45:34 Test Time 10:42:49
Dry Gas 2 0.078 ✓ Time 10:46:34 Report Y

Diagnostic Test Record
Serial Number 130134 Model DataMaster
Location HNH1 Hoonah Police Dept.

Diagnostic Test Results
Test Date 08/28/2009 ✓ Computer Okay
Test Time 10:40:26 Program Okay
Software Date 01/08/2009
Sample Chamber (C) 49
Breath Tube (C) 42
Barometer (in) 29.47
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay